

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041124\
 Data File : VU058636.D
 Acq On : 12 Apr 2024 02:38
 Operator : MD/SY
 Sample : P2055-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COBN7

Quant Time: Apr 12 05:02:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 12 04:41:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	74746	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	73653	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	31451	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	19880	4.017	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.400%
7) Chloroethane-d5	1.907	69	20028	4.657	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.200%
11) 1,1-Dichloroethene-d2	2.560	65	9673	4.319	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.400%
20) 2-Butanone-d5	4.621	46	51884	49.751	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.500%
24) Chloroform-d	5.052	84	46609	4.718	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
26) 1,2-Dichloroethane-d4	5.695	65	23350	5.065	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	5.717	84	89977	4.538	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
36) 1,2-Dichloropropane-d6	6.682	67	29809	4.794	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.800%
41) Toluene-d8	7.891	98	73589	4.185	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
43) trans-1,3-Dichloroprop...	8.174	79	8075	4.183	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.600%
46) 2-Hexanone-d5	8.631	63	33196	41.695	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.380%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	19044	4.338	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	23076	4.356	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.200%
Target Compounds						
					Qvalue	
15) Methyl Acetate	2.959	43	2189	1.042	ug/L	# 67
47) Tetrachloroethene	8.547	164	11801	2.392	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041124\
Data File : VU058636.D
Acq On : 12 Apr 2024 02:38
Operator : MD/SY
Sample : P2055-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BN7

Quant Time: Apr 12 05:02:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 12 04:41:05 2024
Response via : Initial Calibration

