

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041219\
 Data File : VU031130.D
 Acq On : 12 Apr 2019 15:55
 Operator : JC/SP
 Sample : K2299-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Apr 13 01:44:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 01:01:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	437584	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	438506	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	183238	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	191358	48.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.98%
7) Chloroethane-d5	1.68	69	147394	49.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.44%
11) 1,1-Dichloroethene-d2	2.27	63	258205	36.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.10%
21) 2-Butanone-d5	4.17	46	327381	105.75	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.75%
24) Chloroform-d	4.64	84	304316	51.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.00%
26) 1,2-Dichloroethane-d4	5.31	65	215696	52.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.96%
32) Benzene-d6	5.34	84	635336	49.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.26%
36) 1,2-Dichloropropane-d6	6.32	67	219784	50.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.20%
41) Toluene-d8	7.56	98	536127	47.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.50%
43) trans-1,3-Dichloropropene-	7.85	79	94586	47.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.56%
47) 2-Hexanone-d5	8.31	63	201460	100.47	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.47%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	298246	50.40	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	197743	54.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.74%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041219\
Data File : VU031130.D
Acq On : 12 Apr 2019 15:55
Operator : JC/SP
Sample : K2299-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Apr 13 01:44:38 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 01:01:33 2019
Response via : Initial Calibration

