

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041219\
 Data File : VU031136.D
 Acq On : 12 Apr 2019 18:13
 Operator : JC/SP
 Sample : K2353-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK02

Quant Time: Apr 13 02:15:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 01:01:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	468650	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	463591	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	189908	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	208801	49.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.82%
7) Chloroethane-d5	1.68	69	158105	49.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.58%
11) 1,1-Dichloroethene-d2	2.27	63	271079	35.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.66%
21) 2-Butanone-d5	4.17	46	321883	97.08	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.08%
24) Chloroform-d	4.64	84	319241	49.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.90%
26) 1,2-Dichloroethane-d4	5.30	65	218850	50.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.38%
32) Benzene-d6	5.34	84	667151	49.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	6.32	67	223818	48.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.52%
41) Toluene-d8	7.56	98	578132	48.19	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.38%
43) trans-1,3-Dichloropropene-	7.85	79	93752	44.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.66%
47) 2-Hexanone-d5	8.31	63	196237	92.57	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.57%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	302061	48.28	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.56%
64) 1,2-Dichlorobenzene-d4	11.86	152	202274	54.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.32%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041219\
Data File : VU031136.D
Acq On : 12 Apr 2019 18:13
Operator : JC/SP
Sample : K2353-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK02

Quant Time: Apr 13 02:15:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 01:01:33 2019
Response via : Initial Calibration

