

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU041221\  
 Data File : VU043025.D  
 Acq On : 12 Apr 2021 16:01  
 Operator : SY/MD  
 Sample : M1964-02  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 TB-4521

Quant Time: Apr 13 05:13:45 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U041221W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Apr 12 13:17:29 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|--------|------|----------|-------|---------|----------|
| -----                       |        |      |          |       |         |          |
| Internal Standards          |        |      |          |       |         |          |
| 1) Pentafluorobenzene       | 5.382  | 168  | 125169   | 50.00 | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.256  | 114  | 206081   | 50.00 | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 9.423  | 117  | 202067   | 50.00 | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 11.819 | 152  | 100722   | 50.00 | ug/l    | 0.00     |
| System Monitoring Compounds |        |      |          |       |         |          |
| 33) 1,2-Dichloroethane-d4   | 5.710  | 65   | 105455   | 50.12 | ug/l    | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 100.24% |          |
| 35) Dibromofluoromethane    | 5.298  | 113  | 72734    | 46.97 | ug/l    | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 93.94%  |          |
| 50) Toluene-d8              | 7.906  | 98   | 266866   | 48.70 | ug/l    | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 97.40%  |          |
| 62) 4-Bromofluorobenzene    | 10.639 | 95   | 98373    | 44.83 | ug/l    | 0.00     |
| Spiked Amount               | 50.000 |      | Recovery | =     | 89.66%  |          |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU041221\  
Data File : VU043025.D  
Acq On : 12 Apr 2021 16:01  
Operator : SY/MD  
Sample : M1964-02  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
TB-4521

Quant Time: Apr 13 05:13:45 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U041221W.M  
Quant Title : SW846 8260  
QLast Update : Mon Apr 12 13:17:29 2021  
Response via : Initial Calibration

