

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041221\
 Data File : VU043026.D
 Acq On : 12 Apr 2021 16:24
 Operator : SY/MD
 Sample : M1966-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB 2021-04-06

Quant Time: Apr 13 05:14:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	122564	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	202210	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	192010	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	96138	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	101987	49.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.00%	
35) Dibromofluoromethane	5.298	113	71438	47.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.04%	
50) Toluene-d8	7.906	98	255558	47.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.06%	
62) 4-Bromofluorobenzene	10.639	95	92121	42.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041221\
Data File : VU043026.D
Acq On : 12 Apr 2021 16:24
Operator : SY/MD
Sample : M1966-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB 2021-04-06

Quant Time: Apr 13 05:14:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

