

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041221\
 Data File : VU043030.D
 Acq On : 12 Apr 2021 17:56
 Operator : SY/MD
 Sample : M1967-13
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB 2021-04-06 SED

Quant Time: Apr 13 05:15:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	117103	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	192585	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	187618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	93966	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	97758	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
35) Dibromofluoromethane	5.298	113	68742	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
50) Toluene-d8	7.906	98	247129	48.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.52%	
62) 4-Bromofluorobenzene	10.639	95	90639	44.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.40%	
Target Compounds						
16) Acetone	2.629	43	3400	2.63	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041221\
Data File : VU043030.D
Acq On : 12 Apr 2021 17:56
Operator : SY/MD
Sample : M1967-13
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB 2021-04-06_SED

Quant Time: Apr 13 05:15:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

