


Chloromethane

Response Ratio



$$\text{Response} = 5.365\text{e-}001 * \text{Amt} + 1.323\text{e-}002$$


Coef of Det (r^2) = 0.997755 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M

Calibration Table Last Updated: Mon Apr 12 13:17:29 2021

Chloroethane

Response Ratio



$$\text{Response} = 3.878\text{e-}001 * \text{Amt} + 3.006\text{e-}004$$


Coef of Det (r^2) = 0.999703 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M

Calibration Table Last Updated: Mon Apr 12 13:17:29 2021

Acrolein

Response Ratio



$$\text{Response} = 1.409\text{e-}001 * \text{Amt} - 2.936\text{e-}002$$


Coef of Det (r^2) = 0.999837 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M

Calibration Table Last Updated: Mon Apr 12 13:17:29 2021

Acetone

Response Ratio



$$\text{Response} = 4.941\text{e-}001 * \text{Amt} + 2.997\text{e-}003$$


Coef of Det (r^2) = 0.999611 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M

Calibration Table Last Updated: Mon Apr 12 13:17:29 2021

2-Chloroethyl Vinyl ether


Response Ratio



Response = 2.827e-001 * Amt - 4.125e-002
 Coef of Det (r^2) = 0.990778 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Calibration Table Last Updated: Mon Apr 12 13:17:29 2021

1,2,4-Trichlorobenzene

Response Ratio



$$\text{Response} = 1.180\text{e}+000 * \text{Amt} - 3.556\text{e}-002$$


Coef of Det (r^2) = 0.998986 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M

Calibration Table Last Updated: Mon Apr 12 13:17:29 2021

Hexachlorobutadiene

Response Ratio



$$\text{Response} = 5.667\text{e-}001 * \text{Amt} + 6.247\text{e-}003$$


Coef of Det (r^2) = 0.999585 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M

Calibration Table Last Updated: Mon Apr 12 13:17:29 2021

Naphthalene

Response Ratio



$$\text{Response} = 3.827\text{e}+000 * \text{Amt} - 2.098\text{e}-001$$


Coef of Det (r^2) = 0.998192 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M

Calibration Table Last Updated: Mon Apr 12 13:17:29 2021

1,2,3-Trichlorobenzene

Response Ratio



$$\text{Response} = 1.143\text{e}+000 * \text{Amt} - 3.338\text{e}-002$$

Coef of Det (r^2) = 0.999136 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M

Calibration Table Last Updated: Mon Apr 12 13:17:29 2021