

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041222\
 Data File : VU047972.D
 Acq On : 12 Apr 2022 21:43
 Operator : SY/MD
 Sample : VU0412WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK087

Quant Time: Apr 13 04:41:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 13 04:40:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	346388	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	365350	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	192734	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	130919	52.120	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.240%
7) Chloroethane-d5	1.915	69	109749	55.411	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	110.820%
11) 1,1-Dichloroethene-d2	2.568	63	166608	36.253	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.500%
21) 2-Butanone-d5	4.620	46	203992	94.616	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.620%
24) Chloroform-d	5.066	84	228349	49.683	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.360%
26) 1,2-Dichloroethane-d4	5.703	65	146804	52.079	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.160%
32) Benzene-d6	5.729	84	470931	51.226	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.460%
36) 1,2-Dichloropropane-d6	6.693	67	149472	52.942	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.880%
41) Toluene-d8	7.899	98	425232	47.475	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.960%
43) trans-1,3-Dichloroprop...	8.179	79	62814	43.915	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.840%
47) 2-Hexanone-d5	8.632	63	144802	87.611	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.610%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	222130	45.844	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.680%
66) 1,2-Dichlorobenzene-d4	12.195	152	175666	51.231	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.460%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041222\
Data File : VU047972.D
Acq On : 12 Apr 2022 21:43
Operator : SY/MD
Sample : VU0412WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK087

Quant Time: Apr 13 04:41:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 13 04:40:51 2022
Response via : Initial Calibration

