

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041224\
 Data File : VU058646.D
 Acq On : 12 Apr 2024 12:05
 Operator : MD/SY
 Sample : P2013-21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E0MC5

Quant Time: Apr 13 00:04:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 12 23:43:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	71243	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	70926	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	29061	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18981	4.024	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.400%
7) Chloroethane-d5	1.907	69	18755	4.575	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.600%
11) 1,1-Dichloroethene-d2	2.560	65	9698	4.543	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.800%
20) 2-Butanone-d5	4.621	46	51655	51.967	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.940%
24) Chloroform-d	5.052	84	46335	4.921	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.400%
26) 1,2-Dichloroethane-d4	5.692	65	23512	5.351	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.000%
32) Benzene-d6	5.717	84	88375	4.628	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
36) 1,2-Dichloropropane-d6	6.682	67	29133	4.865	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.200%
41) Toluene-d8	7.891	98	71581	4.227	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	8.177	79	8342	4.488	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.800%
46) 2-Hexanone-d5	8.630	63	34966	45.606	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.220%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	19132	4.526	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	22490	4.595	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041224\
Data File : VU058646.D
Acq On : 12 Apr 2024 12:05
Operator : MD/SY
Sample : P2013-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E0MC5

Quant Time: Apr 13 00:04:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 12 23:43:07 2024
Response via : Initial Calibration

