

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041224\
 Data File : VU058670.D
 Acq On : 12 Apr 2024 22:07
 Operator : MD/SY
 Sample : P2035-01DL 4X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AN7DL

Quant Time: Apr 13 01:05:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 12 23:43:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	64280	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	61161	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	25784	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.593	65	16198	3.806	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.200%	
7) Chloroethane-d5	1.904	69	17357	4.693	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.800%	
11) 1,1-Dichloroethene-d2	2.560	65	8121	4.216	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	84.400%	
20) 2-Butanone-d5	4.621	46	51126	57.006	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	114.020%	
24) Chloroform-d	5.052	84	44584	5.248	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.000%	
26) 1,2-Dichloroethane-d4	5.695	65	21472	5.416	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.400%	
32) Benzene-d6	5.718	84	77696	4.719	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.400%	
36) 1,2-Dichloropropane-d6	6.682	67	27141	5.256	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.200%	
41) Toluene-d8	7.891	98	61459	4.209	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.200%	
43) trans-1,3-Dichloroprop...	8.177	79	7657	4.777	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	95.600%	
46) 2-Hexanone-d5	8.631	63	34510	52.198	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	104.400%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	18478	5.069	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.400%	
66) 1,2-Dichlorobenzene-d4	12.190	152	21216	4.885	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.800%	
Target Compounds						Qvalue
16) Methylene chloride	3.039	84	1773	0.293	ug/L	83
22) cis-1,2-Dichloroethene	4.660	96	3462	0.646	ug/L	93
34) Trichloroethene	6.534	95	39255	7.024	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041224\
Data File : VU058670.D
Acq On : 12 Apr 2024 22:07
Operator : MD/SY
Sample : P2035-01DL 4X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AN7DL

Quant Time: Apr 13 01:05:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 12 23:43:07 2024
Response via : Initial Calibration

