

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041224\
 Data File : VU058671.D
 Acq On : 12 Apr 2024 22:30
 Operator : MD/SY
 Sample : P2035-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Apr 13 01:05:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 12 23:43:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	68540	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	67946	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	28361	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	16656	3.671	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	73.400%	
7) Chloroethane-d5	1.907	69	17278	4.381	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	87.600%	
11) 1,1-Dichloroethene-d2	2.557	65	8359	4.070	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	81.400%	
20) 2-Butanone-d5	4.621	46	50347	52.648	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	105.300%	
24) Chloroform-d	5.052	84	44487	4.911	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.200%	
26) 1,2-Dichloroethane-d4	5.695	65	22047	5.215	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.400%	
32) Benzene-d6	5.718	84	80658	4.409	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.200%	
36) 1,2-Dichloropropane-d6	6.682	67	28178	4.912	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.200%	
41) Toluene-d8	7.891	98	64275	3.962	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.200%	
43) trans-1,3-Dichloroprop...	8.177	79	7673	4.309	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.200%	
46) 2-Hexanone-d5	8.631	63	33823	46.050	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	92.100%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	18634	4.601	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	92.000%	
66) 1,2-Dichlorobenzene-d4	12.190	152	21109	4.419	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	88.400%	
Target Compounds						Qvalue
16) Methylene chloride	3.036	84	1431	0.222	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041224\
Data File : VU058671.D
Acq On : 12 Apr 2024 22:30
Operator : MD/SY
Sample : P2035-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Apr 13 01:05:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 12 23:43:07 2024
Response via : Initial Calibration

