




Data File : VU031191.D
Acq On : 14 Apr 2019 05:32
Operator : JC/SP
Sample : K2409-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Quant Time: Apr 15 01:43:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Apr 15 01:39:53 2019
Response via : Initial Calibration



TIC: VU031191.D

(30) 1,1,1-Trichloroethane (T)

4.910min (+0.003) 0.57ug/L

response 2752

Ion	Exp%	Act%
97.00	100	100
99.00	64.30	93.02#
61.05	46.40	61.41#
0.00	0.00	0.00