

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041319\
 Data File : VU031154.D
 Acq On : 13 Apr 2019 14:18
 Operator : JC/SP
 Sample : VU0413WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK60

Quant Time: Apr 15 11:36:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 13 22:25:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	414202	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	385541	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	143870	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	177543	48.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.04%
7) Chloroethane-d5	1.67	69	141424	50.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.80%
11) 1,1-Dichloroethene-d2	2.26	63	244697	36.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.18%
21) 2-Butanone-d5	4.17	46	286958	97.92	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.92%
24) Chloroform-d	4.64	84	295684	52.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.70%
26) 1,2-Dichloroethane-d4	5.30	65	203625	52.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.66%
32) Benzene-d6	5.33	84	590187	52.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.88%
36) 1,2-Dichloropropane-d6	6.32	67	204981	53.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.28%
41) Toluene-d8	7.56	98	457470	45.86	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.72%
43) trans-1,3-Dichloropropene-	7.85	79	81089	46.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.20%
47) 2-Hexanone-d5	8.31	63	158940	90.16	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.16%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	272882	52.45	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.90%
64) 1,2-Dichlorobenzene-d4	11.86	152	161657	57.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.26%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041319\
Data File : VU031154.D
Acq On : 13 Apr 2019 14:18
Operator : JC/SP
Sample : VU0413WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK60

Quant Time: Apr 15 11:36:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 13 22:25:59 2019
Response via : Initial Calibration

