

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
 Data File : VU043055.D
 Acq On : 13 Apr 2021 16:36
 Operator : SY/MD
 Sample : PB135653TB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB135653TB

Quant Time: Apr 14 03:51:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	159385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	269886	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	256834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	130533	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	159895	59.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.36%	
35) Dibromofluoromethane	5.298	113	110402	54.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.88%	
50) Toluene-d8	7.906	98	401765	55.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.98%	
62) 4-Bromofluorobenzene	10.639	95	149645	52.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.16%	
Target Compounds						
16) Acetone	2.639	43	15572	9.58	ug/l	99
20) Methylene Chloride	3.054	84	4727	2.48	ug/l	97
43) Isopropyl Acetate	5.919	43	21997	3.15	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
Data File : VU043055.D
Acq On : 13 Apr 2021 16:36
Operator : SY/MD
Sample : PB135653TB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB135653TB

Quant Time: Apr 14 03:51:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

