

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
 Data File : VU043056.D
 Acq On : 13 Apr 2021 16:59
 Operator : SY/MD
 Sample : PB135653ZHE#01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB135653ZHE#01

Quant Time: Apr 14 03:52:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	155638	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	263698	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	257836	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	135626	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	156994	60.00	ug/l	0.00
Spiked Amount 50.000			Recovery	= 120.00%		
35) Dibromofluoromethane	5.298	113	108183	54.60	ug/l	0.00
Spiked Amount 50.000			Recovery	= 109.20%		
50) Toluene-d8	7.906	98	397648	56.71	ug/l	0.00
Spiked Amount 50.000			Recovery	= 113.42%		
62) 4-Bromofluorobenzene	10.639	95	152952	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	= 108.96%		
Target Compounds						
						Qvalue
16) Acetone	2.639	43	15728	9.92	ug/l	99
18) Methyl Acetate	2.961	43	4677	1.43	ug/l	93
20) Methylene Chloride	3.054	84	3951	2.13	ug/l #	89
43) Isopropyl Acetate	5.919	43	21559	3.16	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
Data File : VU043056.D
Acq On : 13 Apr 2021 16:59
Operator : SY/MD
Sample : PB135653ZHE#01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB135653ZHE#01

Quant Time: Apr 14 03:52:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

