

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
 Data File : VU043057.D
 Acq On : 13 Apr 2021 17:22
 Operator : SY/MD
 Sample : PB135653ZHE#02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB135653ZHE#02

Quant Time: Apr 14 03:52:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	160837	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	267078	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	263183	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	136714	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	161391	59.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.38%	
35) Dibromofluoromethane	5.298	113	110939	55.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.56%	
50) Toluene-d8	7.906	98	408313	57.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.00%	
62) 4-Bromofluorobenzene	10.639	95	157755	55.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.96%	
Target Compounds						
						Qvalue
16) Acetone	2.639	43	15138	9.22	ug/l	95
18) Methyl Acetate	2.961	43	3922	1.16	ug/l	92
20) Methylene Chloride	3.051	84	4244	2.21	ug/l	93
43) Isopropyl Acetate	5.919	43	20681	2.99	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
Data File : VU043057.D
Acq On : 13 Apr 2021 17:22
Operator : SY/MD
Sample : PB135653ZHE#02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB135653ZHE#02

Quant Time: Apr 14 03:52:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

