

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
 Data File : VU043058.D
 Acq On : 13 Apr 2021 17:45
 Operator : SY/MD
 Sample : PB135653ZHE#03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB135653ZHE#03

Quant Time: Apr 14 03:52:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	151672	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	251335	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	255118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	128608	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	151004	59.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.44%	
35) Dibromofluoromethane	5.298	113	103685	54.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.80%	
50) Toluene-d8	7.906	98	390827	58.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.96%	
62) 4-Bromofluorobenzene	10.639	95	150074	56.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.16%	
Target Compounds						
16) Acetone	2.639	43	15266	9.88	ug/l	98
18) Methyl Acetate	2.957	43	4760	1.49	ug/l	100
20) Methylene Chloride	3.051	84	4259	2.35	ug/l	94
43) Isopropyl Acetate	5.919	43	20687	3.18	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
Data File : VU043058.D
Acq On : 13 Apr 2021 17:45
Operator : SY/MD
Sample : PB135653ZHE#03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB135653ZHE#03

Quant Time: Apr 14 03:52:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

