

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
 Data File : VU043059.D
 Acq On : 13 Apr 2021 18:08
 Operator : SY/MD
 Sample : M1998-14
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB_04-08-2021_TB

Quant Time: Apr 14 03:52:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	147628	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	243709	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	232754	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	120774	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	121270	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
35) Dibromofluoromethane	5.298	113	84541	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
50) Toluene-d8	7.906	98	307645	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
62) 4-Bromofluorobenzene	10.639	95	116019	44.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.42%	
Target Compounds						
18) Methyl Acetate	2.964	43	3245	1.05	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
Data File : VU043059.D
Acq On : 13 Apr 2021 18:08
Operator : SY/MD
Sample : M1998-14
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB_04-08-2021_TB

Quant Time: Apr 14 03:52:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

