

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
 Data File : VU043060.D
 Acq On : 13 Apr 2021 18:31
 Operator : SY/MD
 Sample : M1998-13
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB_04-08-2021_SO

Quant Time: Apr 14 03:53:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	146091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	239893	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	232100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	118817	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	119486	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
35) Dibromofluoromethane	5.298	113	84580	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
50) Toluene-d8	7.906	98	310239	48.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.28%	
62) 4-Bromofluorobenzene	10.639	95	115566	45.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.48%	
Target Compounds						
					Qvalue	
16) Acetone	2.642	43	4212	2.61	ug/l #	89
18) Methyl Acetate	2.961	43	3209	1.05	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
Data File : VU043060.D
Acq On : 13 Apr 2021 18:31
Operator : SY/MD
Sample : M1998-13
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB_04-08-2021_SO

Quant Time: Apr 14 03:53:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

