

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
 Data File : VU043071.D
 Acq On : 13 Apr 2021 23:29
 Operator : SY/MD
 Sample : M1964-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FS-SB16

Quant Time: Apr 14 05:09:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	137309	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	225653	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	211444	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	106025	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	111666	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
35) Dibromofluoromethane	5.298	113	78942	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
50) Toluene-d8	7.906	98	278234	46.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.74%	
62) 4-Bromofluorobenzene	10.639	95	104080	43.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.64%	
Target Compounds						Qvalue
11) Tert butyl alcohol	3.192	59	2452	4.12	ug/l	99
16) Acetone	2.633	43	9777	6.90	ug/l	96
20) Methylene Chloride	3.057	84	1835	1.12	ug/l	98
25) 2-Butanone	4.719	43	3003	1.43	ug/l	91
27) cis-1,2-Dichloroethene	4.674	96	787	0.46	ug/l	40
30) Chloroform	5.096	83	38464	12.04	ug/l	90
40) Benzene	5.784	78	1780	0.26	ug/l #	90
44) Trichloroethene	6.549	130	1140	0.66	ug/l	81
52) Toluene	7.977	92	1621	0.38	ug/l	96
95) Naphthalene	14.095	128	7600	3.68	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
Data File : VU043071.D
Acq On : 13 Apr 2021 23:29
Operator : SY/MD
Sample : M1964-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FS-SB16

Quant Time: Apr 14 05:09:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

