

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
 Data File : VU043076.D
 Acq On : 14 Apr 2021 01:24
 Operator : SY/MD
 Sample : M1966-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-122

Quant Time: Apr 14 06:00:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	130708	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	215431	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	203123	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	100518	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	105204	47.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.76%	
35) Dibromofluoromethane	5.298	113	75497	46.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.28%	
50) Toluene-d8	7.906	98	269510	47.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.10%	
62) 4-Bromofluorobenzene	10.639	95	99924	43.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.12%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	3.195	59	29533	52.15	ug/l	99
16) Acetone	2.633	43	8285	6.11	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041321\
Data File : VU043076.D
Acq On : 14 Apr 2021 01:24
Operator : SY/MD
Sample : M1966-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-122

Quant Time: Apr 14 06:00:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

