

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
 Data File : VU048011.D
 Acq On : 13 Apr 2022 18:07
 Operator : SY/MD
 Sample : N2316-10ME
 Misc : 6.49g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETNQ9ME

Quant Time: Apr 14 03:10:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 14 02:49:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	404649	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	434954	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	253911	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	173605	59.163	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	118.320%
7) Chloroethane-d5	1.871	69	84022	36.314	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	72.620%
11) 1,1-Dichloroethene-d2	2.536	63	229124	42.678	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.360%
21) 2-Butanone-d5	4.632	46	340248	135.092	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	135.090%#
24) Chloroform-d	5.063	84	347160	64.659	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	129.320%#
26) 1,2-Dichloroethane-d4	5.700	65	227607	69.119	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	138.240%#
32) Benzene-d6	5.722	84	725785	66.314	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	132.620%#
36) 1,2-Dichloropropane-d6	6.690	67	239955	71.390	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	142.780%#
41) Toluene-d8	7.899	98	677747	63.559	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	127.120%#
43) trans-1,3-Dichloroprop...	8.182	79	108022	63.436	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	126.880%#
47) 2-Hexanone-d5	8.648	63	288640	146.693	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	146.690%#
56) 1,1,2,2-Tetrachloroeth...	10.764	84	404459	70.116	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	140.240%#
66) 1,2-Dichlorobenzene-d4	12.198	152	313661	69.435	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	138.880%#
Target Compounds						
					Qvalue	
33) Benzene	5.771	78	46339	3.681	ug/L	100
51) Chlorobenzene	9.452	112	184474	19.668	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
Data File : VU048011.D
Acq On : 13 Apr 2022 18:07
Operator : SY/MD
Sample : N2316-10ME
Misc : 6.49g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETNQ9ME

Quant Time: Apr 14 03:10:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 14 02:49:33 2022
Response via : Initial Calibration

