

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
 Data File : VU048018.D
 Acq On : 13 Apr 2022 21:12
 Operator : SY/MD
 Sample : VU0413WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK090

Quant Time: Apr 14 06:53:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 14 06:53:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	345611	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	358988	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	190909	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	122891	49.034	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.060%
7) Chloroethane-d5	1.916	69	104000	52.627	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.260%
11) 1,1-Dichloroethene-d2	2.568	63	159974	34.888	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.780%
21) 2-Butanone-d5	4.620	46	203801	94.740	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.740%
24) Chloroform-d	5.067	84	224405	48.935	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.880%
26) 1,2-Dichloroethane-d4	5.703	65	146578	52.116	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.240%
32) Benzene-d6	5.729	84	461752	51.117	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.240%
36) 1,2-Dichloropropane-d6	6.690	67	148024	53.358	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.720%
41) Toluene-d8	7.899	98	414888	47.141	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.280%
43) trans-1,3-Dichloroprop...	8.179	79	62241	44.286	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.580%
47) 2-Hexanone-d5	8.632	63	150715	92.805	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.810%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	223556	46.956	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.920%
66) 1,2-Dichlorobenzene-d4	12.195	152	176073	51.840	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
Data File : VU048018.D
Acq On : 13 Apr 2022 21:12
Operator : SY/MD
Sample : VU0413WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK090

Quant Time: Apr 14 06:53:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 14 06:53:20 2022
Response via : Initial Calibration

