

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
 Data File : VU048034.D
 Acq On : 14 Apr 2022 03:22
 Operator : SY/MD
 Sample : N2387-05 100X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0J57

Quant Time: Apr 14 07:20:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 14 06:53:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	328411	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	345865	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	180694	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	102047	42.850	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.700%
7) Chloroethane-d5	1.916	69	93036	49.544	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.080%
11) 1,1-Dichloroethene-d2	2.571	63	140742	32.301	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.600%
21) 2-Butanone-d5	4.620	46	190690	93.287	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.290%
24) Chloroform-d	5.067	84	206709	47.437	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.880%
26) 1,2-Dichloroethane-d4	5.703	65	134594	50.361	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.720%
32) Benzene-d6	5.729	84	417591	47.983	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.960%
36) 1,2-Dichloropropane-d6	6.693	67	135495	50.695	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.400%
41) Toluene-d8	7.899	98	368007	43.401	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	8.179	79	54388	40.166	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.340%
47) 2-Hexanone-d5	8.632	63	136194	87.046	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.050%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	208481	45.451	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.900%
66) 1,2-Dichlorobenzene-d4	12.195	152	160505	49.928	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
Data File : VU048034.D
Acq On : 14 Apr 2022 03:22
Operator : SY/MD
Sample : N2387-05 100X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0J57

Quant Time: Apr 14 07:20:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 14 06:53:20 2022
Response via : Initial Calibration

