




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
 Data File : VU048015.D
 Acq On : 13 Apr 2022 19:40
 Operator : SY/MD
 Sample : N2316-04ME
 Misc : 6.86g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETNQ2ME

Manual IntegrationsAPPROVED

Quant Time: Apr 14 03:11:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 14 02:49:33 2022
 Response via : Initial Calibration

Reviewed By :John Carlone 04/14/2022
 Supervised By :Mahesh Dadoda 04/14/2022



TIC: VU048015.D\data.ms

(7) Chloroethane-d5 (S)

1.893min (-0.019) 25.09 ug/L

response 56376

Ion	Exp%	Act%
69.10	100.00	100.00
71.10	32.10	33.94
51.00	24.50	20.61
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
 Data File : VU048015.D
 Acq On : 13 Apr 2022 19:40
 Operator : SY/MD
 Sample : N2316-04ME
 Misc : 6.86g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETNQ2ME

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 04/14/2022
 Supervised By :Mahesh Dadoda 04/14/2022

Quant Time: Apr 14 03:11:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 14 02:49:33 2022
 Response via : Initial Calibration



TIC: VU048015.D\data.ms

(7) Chloroethane-d5 (S)

1.893min (-0.019) 44.44 ug/L m

response 99856

Ion	Exp%	Act%
69.10	100.00	100.00
71.10	32.10	19.16#
51.00	24.50	11.63#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
 Data File : VU048015.D
 Acq On : 13 Apr 2022 19:40
 Operator : SY/MD
 Sample : N2316-04ME
 Misc : 6.86g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETNQ2ME

Manual IntegrationsAPPROVED

Reviewed By : John Carlone 04/14/2022
 Supervised By : Mahesh Dadoda 04/14/2022

Quant Time: Apr 14 03:11:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 14 02:49:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	392939	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	411750	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	227055	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	177841	62.412	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	124.820%	
7) Chloroethane-d5	1.893	69	99856m	44.444	ug/L	-0.02
Spiked Amount 50.000	Range 70	- 130	Recovery	=	88.880%	
11) 1,1-Dichloroethene-d2	2.539	63	244653	46.928	ug/L	-0.03
Spiked Amount 50.000	Range 60	- 125	Recovery	=	93.860%	
21) 2-Butanone-d5	4.642	46	349547	142.920	ug/L	0.02
Spiked Amount 100.000	Range 40	- 130	Recovery	=	142.920%#	
24) Chloroform-d	5.063	84	354887	68.067	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	136.140%#	
26) 1,2-Dichloroethane-d4	5.703	65	233788	73.111	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	146.220%#	
32) Benzene-d6	5.723	84	747356	72.133	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	144.260%#	
36) 1,2-Dichloropropane-d6	6.690	67	242941	76.351	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	152.700%#	
41) Toluene-d8	7.899	98	686685	68.026	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	136.060%#	
43) trans-1,3-Dichloroprop...	8.182	79	103606	64.271	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	128.540%#	
47) 2-Hexanone-d5	8.661	63	280487	150.582	ug/L	0.03
Spiked Amount 100.000	Range 45	- 130	Recovery	=	150.580%#	
56) 1,1,2,2-Tetrachloroeth...	10.767	84	379264	69.454	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	138.900%#	
66) 1,2-Dichlorobenzene-d4	12.198	152	300735	74.448	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	148.900%#	
Target Compounds						
33) Benzene	5.771	78	56789	4.765	ug/L	100
42) Toluene	7.973	91	19227	1.470	ug/L	99
51) Chlorobenzene	9.452	112	1918148	216.026	ug/L	98
65) 1,4-Dichlorobenzene	11.841	146	7498	1.061	ug/L	99
67) 1,2-Dichlorobenzene	12.214	146	5209	0.763	ug/L #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
Data File : VU048015.D
Acq On : 13 Apr 2022 19:40
Operator : SY/MD
Sample : N2316-04ME
Misc : 6.86g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETNQ2ME

Manual IntegrationsAPPROVED

Quant Time: Apr 14 03:11:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 14 02:49:33 2022
Response via : Initial Calibration

Reviewed By :John Carlone 04/14/2022
Supervised By :Mahesh Dadoda 04/14/2022

