

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
 Data File : VU047998.D
 Acq On : 13 Apr 2022 13:08
 Operator : SY/MD
 Sample : N2293-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETNP0

Quant Time: Apr 14 03:07:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 14 02:49:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	362148	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	371074	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	212009	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	128038	48.755	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.500%
7) Chloroethane-d5	1.913	69	109482	52.871	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.740%
11) 1,1-Dichloroethene-d2	2.568	63	164693	34.277	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.560%
21) 2-Butanone-d5	4.623	46	215734	95.707	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.710%
24) Chloroform-d	5.067	84	230615	47.993	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.980%
26) 1,2-Dichloroethane-d4	5.703	65	149776	50.821	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.640%
32) Benzene-d6	5.729	84	483331	51.764	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.520%
36) 1,2-Dichloropropane-d6	6.690	67	154925	54.027	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.060%
41) Toluene-d8	7.899	98	428423	47.094	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.180%
43) trans-1,3-Dichloroprop...	8.179	79	67218	46.269	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.540%
47) 2-Hexanone-d5	8.633	63	162482	96.792	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.790%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	233722	47.493	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.980%
66) 1,2-Dichlorobenzene-d4	12.192	152	193119	51.200	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041322\
Data File : VU047998.D
Acq On : 13 Apr 2022 13:08
Operator : SY/MD
Sample : N2293-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETNP0

Quant Time: Apr 14 03:07:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 14 02:49:33 2022
Response via : Initial Calibration

