

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041421\
 Data File : VU043102.D
 Acq On : 14 Apr 2021 18:13
 Operator : SY/MD
 Sample : M2002-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Apr 15 02:08:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	136781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	223314	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	210897	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	101184	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	110596	48.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.20%	
35) Dibromofluoromethane	5.298	113	76653	45.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.36%	
50) Toluene-d8	7.906	98	282479	47.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.14%	
62) 4-Bromofluorobenzene	10.639	95	103617	43.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.16%	

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041421\
Data File : VU043102.D
Acq On : 14 Apr 2021 18:13
Operator : SY/MD
Sample : M2002-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Apr 15 02:08:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

