

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
 Data File : VU048045.D
 Acq On : 14 Apr 2022 10:15
 Operator : SY/MD
 Sample : VU0414WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK092

Quant Time: Apr 15 02:24:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 15 02:23:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	309300	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	331729	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	170324	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	113537	50.620	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 101.240%		
7) Chloroethane-d5	1.916	69	100045	56.569	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 113.140%		
11) 1,1-Dichloroethene-d2	2.568	63	149627	36.462	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 72.920%		
21) 2-Butanone-d5	4.620	46	195760	101.685	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 101.690%		
24) Chloroform-d	5.067	84	217616	53.026	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 106.060%		
26) 1,2-Dichloroethane-d4	5.703	65	140190	55.696	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 111.400%		
32) Benzene-d6	5.729	84	435107	52.126	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.260%		
36) 1,2-Dichloropropane-d6	6.694	67	143021	55.791	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 111.580%		
41) Toluene-d8	7.899	98	387407	47.636	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 95.280%		
43) trans-1,3-Dichloroprop...	8.179	79	59423	45.755	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 91.500%		
47) 2-Hexanone-d5	8.632	63	139008	92.630	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 92.630%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	218124	49.580	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 99.160%		
66) 1,2-Dichlorobenzene-d4	12.195	152	167660	55.329	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 110.660%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
Data File : VU048045.D
Acq On : 14 Apr 2022 10:15
Operator : SY/MD
Sample : VU0414WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK092

Quant Time: Apr 15 02:24:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 15 02:23:12 2022
Response via : Initial Calibration

