

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
 Data File : VU048072.D
 Acq On : 14 Apr 2022 20:44
 Operator : SY/MD
 Sample : VU0414WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK094

Quant Time: Apr 15 05:05:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 15 05:02:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	320320	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	334889	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	175199	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	108470	46.697	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	93.400%	
7) Chloroethane-d5	1.916	69	97907	53.455	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	106.920%	
11) 1,1-Dichloroethene-d2	2.568	63	143001	33.649	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	67.300%	
21) 2-Butanone-d5	4.620	46	208291	104.472	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	104.470%	
24) Chloroform-d	5.067	84	221157	52.035	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	104.060%	
26) 1,2-Dichloroethane-d4	5.703	65	143963	55.227	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	110.460%	
32) Benzene-d6	5.729	84	440097	52.226	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	104.460%	
36) 1,2-Dichloropropane-d6	6.693	67	144873	55.980	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	111.960%	
41) Toluene-d8	7.899	98	385126	46.909	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	93.820%	
43) trans-1,3-Dichloroprop...	8.179	79	58455	44.585	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	89.160%	
47) 2-Hexanone-d5	8.632	63	145526	96.058	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	96.060%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	228353	51.415	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	102.840%	
66) 1,2-Dichlorobenzene-d4	12.195	152	166449	53.401	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	106.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
Data File : VU048072.D
Acq On : 14 Apr 2022 20:44
Operator : SY/MD
Sample : VU0414WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK094

Quant Time: Apr 15 05:05:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 15 05:02:53 2022
Response via : Initial Calibration

