

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
 Data File : VU048079.D
 Acq On : 14 Apr 2022 23:26
 Operator : SY/MD
 Sample : N2374-04ME
 Misc : 5.32g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFFD9ME

Quant Time: Apr 15 05:28:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 15 05:02:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	160535	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	167538	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	95940	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	94326	81.026	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	162.060%#
7) Chloroethane-d5	1.871	69	50324	54.823	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.640%
11) 1,1-Dichloroethene-d2	2.539	63	135480	63.609	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	127.220%#
21) 2-Butanone-d5	4.632	46	239112	239.301	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	239.300%#
24) Chloroform-d	5.063	84	228920	107.470	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	214.940%#
26) 1,2-Dichloroethane-d4	5.700	65	152420	116.670	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	233.340%#
32) Benzene-d6	5.722	84	454827	107.888	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	215.780%#
36) 1,2-Dichloropropane-d6	6.690	67	153432	118.509	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	237.020%#
41) Toluene-d8	7.899	98	399361	97.231	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	194.460%#
43) trans-1,3-Dichloroprop...	8.182	79	62018	94.552	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	189.100%#
47) 2-Hexanone-d5	8.652	63	173266	228.610	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	228.610%#
56) 1,1,2,2-Tetrachloroeth...	10.764	84	277896	125.071	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	250.140%#
66) 1,2-Dichlorobenzene-d4	12.198	152	189358	110.939	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	221.880%#
Target Compounds						Qvalue
3) Chloromethane	1.517	50	2156	1.515	ug/L	93
42) Toluene	7.970	91	6494	1.220	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
Data File : VU048079.D
Acq On : 14 Apr 2022 23:26
Operator : SY/MD
Sample : N2374-04ME
Misc : 5.32g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFFD9ME

Quant Time: Apr 15 05:28:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 15 05:02:53 2022
Response via : Initial Calibration

