

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
 Data File : VU048046.D
 Acq On : 14 Apr 2022 10:38
 Operator : SY/MD
 Sample : N2358-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Apr 15 02:24:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 15 02:23:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	292409	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	311066	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	159403	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	115242	54.348	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 108.700%		
7) Chloroethane-d5	1.916	69	102646	61.392	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 122.780%		
11) 1,1-Dichloroethene-d2	2.568	63	151388	39.022	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 78.040%		
21) 2-Butanone-d5	4.620	46	210943	115.901	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 115.900%		
24) Chloroform-d	5.067	84	223952	57.722	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 115.440%		
26) 1,2-Dichloroethane-d4	5.703	65	143112	60.141	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 120.280%		
32) Benzene-d6	5.729	84	446681	57.067	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 114.140%		
36) 1,2-Dichloropropane-d6	6.694	67	146310	60.865	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 121.740%#		
41) Toluene-d8	7.899	98	387308	50.787	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 101.580%		
43) trans-1,3-Dichloroprop...	8.179	79	62163	51.044	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 102.080%		
47) 2-Hexanone-d5	8.632	63	147938	105.129	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 105.130%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	229047	55.521	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 111.040%		
66) 1,2-Dichlorobenzene-d4	12.195	152	171123	60.341	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 120.680%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
Data File : VU048046.D
Acq On : 14 Apr 2022 10:38
Operator : SY/MD
Sample : N2358-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Apr 15 02:24:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 15 02:23:12 2022
Response via : Initial Calibration

