

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
 Data File : VU048071.D
 Acq On : 14 Apr 2022 20:21
 Operator : SY/MD
 Sample : VU0414MBL02
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK093

Quant Time: Apr 15 01:22:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 14 06:53:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	304858	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	326113	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	169269	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	91276	41.288	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.580%
7) Chloroethane-d5	1.912	69	86256	49.483	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.960%
11) 1,1-Dichloroethene-d2	2.568	63	126926	31.381	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.760%
21) 2-Butanone-d5	4.623	46	201381	106.129	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.130%
24) Chloroform-d	5.067	84	202460	50.051	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.100%
26) 1,2-Dichloroethane-d4	5.703	65	133300	53.730	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.460%
32) Benzene-d6	5.729	84	398381	48.548	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.100%
36) 1,2-Dichloropropane-d6	6.694	67	134497	53.370	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.740%
41) Toluene-d8	7.899	98	342974	42.899	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.800%
43) trans-1,3-Dichloroprop...	8.182	79	55601	43.549	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.100%
47) 2-Hexanone-d5	8.636	63	142992	96.926	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.930%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	215875	49.914	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.820%
66) 1,2-Dichlorobenzene-d4	12.195	152	154957	51.456	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
Data File : VU048071.D
Acq On : 14 Apr 2022 20:21
Operator : SY/MD
Sample : VU0414MBL02
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK093

Quant Time: Apr 15 01:22:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 14 06:53:20 2022
Response via : Initial Calibration

