

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
 Data File : VU048084.D
 Acq On : 15 Apr 2022 01:21
 Operator : SY/MD
 Sample : N2387-16
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0J97

Quant Time: Apr 15 05:29:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 15 05:02:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	292080	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	287521	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	143492	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	78190	36.916	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.840%
7) Chloroethane-d5	1.916	69	76271	45.669	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.340%
11) 1,1-Dichloroethene-d2	2.568	63	111016	28.648	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.300%#
21) 2-Butanone-d5	4.620	46	188013	103.419	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.420%
24) Chloroform-d	5.067	84	182443	47.076	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.160%
26) 1,2-Dichloroethane-d4	5.703	65	122459	51.520	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.040%
32) Benzene-d6	5.729	84	358829	49.597	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.200%
36) 1,2-Dichloropropane-d6	6.690	67	107228	48.260	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.520%
41) Toluene-d8	7.899	98	302285	42.884	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.760%
43) trans-1,3-Dichloroprop...	8.179	79	46908	41.672	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.340%
47) 2-Hexanone-d5	8.632	63	123216	94.731	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.730%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	200860	52.676	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.360%
66) 1,2-Dichlorobenzene-d4	12.195	152	128719	50.421	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041422\
Data File : VU048084.D
Acq On : 15 Apr 2022 01:21
Operator : SY/MD
Sample : N2387-16
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0J97

Quant Time: Apr 15 05:29:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 15 05:02:53 2022
Response via : Initial Calibration

