

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041521\
 Data File : VU043139.D
 Acq On : 15 Apr 2021 22:58
 Operator : SY/MD
 Sample : M2055-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4478

Manual Integrations
 APPROVED

MMDadoda
 4/19/2021 10:21:12 AM

Quant Time: Apr 16 07:15:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 16 07:14:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	65192	5.00	ug/L	# 0.00
28) Chlorobenzene-d5	9.427	117	64632	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	30989	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	18385	5.55	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	111.00%		
7) Chloroethane-d5	1.919	69	17384	5.57	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	111.40%		
11) 1,1-Dichloroethene-d2	2.572	65	11375	5.93	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	118.60%		
20) 2-Butanone-d5	4.642	46	96514m	66.38	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	132.76%#		
24) Chloroform-d	5.073	84	48013	5.70	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	114.00%		
26) 1,2-Dichloroethane-d4	5.710	65	32657	5.73	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	114.60%		
32) Benzene-d6	5.735	84	87706	5.83	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	116.60%		
36) 1,2-Dichloropropane-d6	6.700	67	29624	5.94	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	118.80%		
41) Toluene-d8	7.906	98	80450	5.14	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	102.80%		
43) trans-1,3-Dichloroprop...	8.189	79	11557	5.24	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	104.80%		
46) 2-Hexanone-d5	8.645	63	69042	62.81	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	125.62%		
56) 1,1,2,2-Tetrachloroeth...	10.764	84	26183	5.96	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	119.20%		
66) 1,2-Dichlorobenzene-d4	12.201	152	32550	6.13	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	122.60%#		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.392	85	910	0.22	ug/L	90
13) Acetone	2.658	43	2718m	2.54	ug/L	
17) Methyl tert-butyl Ether	3.372	73	89530	7.56	ug/L	# 93
19) 1,1-Dichloroethane	3.880	63	879	0.10	ug/L	# 82
22) cis-1,2-Dichloroethene	4.678	96	3555	0.78	ug/L	98
25) Chloroform	5.102	83	1816	0.20	ug/L	82
34) Trichloroethene	6.552	95	12408	2.61	ug/L	95
47) Tetrachloroethene	8.562	164	69818	18.50	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041521\
Data File : VU043139.D
Acq On : 15 Apr 2021 22:58
Operator : SY/MD
Sample : M2055-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4478

Manual Integrations
APPROVED

MMDadoda
4/19/2021 10:21:12 AM

Quant Time: Apr 16 07:15:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR041521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 16 07:14:05 2021
Response via : Initial Calibration

