

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041619\
 Data File : VU031207.D
 Acq On : 16 Apr 2019 11:05
 Operator : JC/SP
 Sample : K2136-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 38023-122917

Quant Time: Apr 17 03:12:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U041519DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Apr 16 04:55:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.74	96	37048	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	14956	1.10	ug/l	0.00
Spiked Amount 1.000			Recovery	=	110.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	14386	1.06	ug/l	0.00
Spiked Amount 1.000			Recovery	=	106.00%	
Target Compounds						Ovalue
3) Chloromethane	1.32	50	585330	34.623	ug/l	99
5) Bromomethane	1.63	94	260779	26.951	ug/l	99
6) Chloroethane	1.70	64	144282	12.651	ug/l	99
7) Trichlorofluoromethane	1.89	101	353688	14.685	ug/l	97
10) Iodomethane	2.41	142	52621	3.253	ug/l	99
13) Acetone	2.32	43	3272	1.120	ug/l	99
15) Methylene Chloride	2.70	84	3658	0.237	ug/l	93
17) 1,1-Dichloroethane	3.44	63	296314	14.153	ug/l	100
21) 2,2-Dichloropropane	4.22	77	205693	12.675	ug/l	96
25) Methyl tert-Butyl Ether	3.00	73	964380	27.614	ug/l	100
26) Bromochloromethane	4.55	128	22793	4.338	ug/l	98
29) 1,1-Dichloropropene	5.13	75	163175	11.051	ug/l	100
37) Trichloroethene	6.18	130	4584	0.372	ug/l	92
43) Dibromomethane	6.56	93	57280	9.097	ug/l	98
50) t-1,3-Dichloropropene	7.88	75	229645	17.749	ug/l	99
51) cis-1,3-Dichloropropene	7.27	75	61518	4.022	ug/l	97
53) 1,3-Dichloropropane	8.24	76	148699	10.310	ug/l	98
60) 1,1,1,2-Tetrachloroethane	9.21	131	73416	7.312	ug/l	99
61) Pentachloroethane	11.10	117	2122	0.272	ug/l	89
62) Hexachloroethane	12.14	117	2621	0.347	ug/l	98
69) Isopropylbenzene	10.16	105	329778	7.898	ug/l	100
70) 1,1,2,2-Tetrachloroethane	10.45	83	163672	14.853	ug/l	99
71) 1,2,3-Trichloropropane	10.49	75	124130	15.579	ug/l #	99
73) n-propylbenzene	10.58	120	57646	5.170	ug/l	98
76) 4-Chlorotoluene	10.77	126	128221	12.378	ug/l	83
78) 1,2,4-Trimethylbenzene	11.14	105	194039	5.447	ug/l	100
79) sec-Butylbenzene	11.32	105	428115	9.203	ug/l	99
81) p-Isopropyltoluene	11.47	119	368142	9.852	ug/l	99
82) 1,3-Dichlorobenzene	11.41	146	385619	17.671	ug/l	99
84) n-Butylbenzene	11.89	91	241256	6.681	ug/l	99
88) Hexachlorobutadiene	13.69	225	296361	35.475	ug/l	98
90) 1,2,3-Trichlorobenzene	13.98	180	103416	8.259	ug/l	99
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041619\
Data File : VU031207.D
Acq On : 16 Apr 2019 11:05
Operator : JC/SP
Sample : K2136-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
38023-122917

Quant Time: Apr 17 03:12:08 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U041519DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Apr 16 04:55:56 2019
Response via : Initial Calibration

