

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041621\
 Data File : VU043164.D
 Acq On : 16 Apr 2021 13:33
 Operator : SY/MD
 Sample : M2061-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB-01-041521

Quant Time: Apr 17 02:30:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	111585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	186834	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	177963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	83925	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	96579	51.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.98%	
35) Dibromofluoromethane	5.298	113	66397	47.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.58%	
50) Toluene-d8	7.906	98	237246	47.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.52%	
62) 4-Bromofluorobenzene	10.639	95	85792	43.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.26%	
Target Compounds						
						Qvalue
16) Acetone	2.642	43	51449	46.36	ug/l	97
20) Methylene Chloride	3.051	84	1827	1.37	ug/l #	86
25) 2-Butanone	4.729	43	12025	7.06	ug/l	91
51) 4-Methyl-2-Pentanone	7.803	43	7211	2.02	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041621\
Data File : VU043164.D
Acq On : 16 Apr 2021 13:33
Operator : SY/MD
Sample : M2061-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB-01-041521

Quant Time: Apr 17 02:30:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

