

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041621\
 Data File : VU043166.D
 Acq On : 16 Apr 2021 14:19
 Operator : SY/MD
 Sample : M2061-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GWBR1-225-250-041521

Quant Time: Apr 17 02:31:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	112885	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	185130	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	179522	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	84954	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	96728	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
35) Dibromofluoromethane	5.298	113	67754	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
50) Toluene-d8	7.906	98	238724	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
62) 4-Bromofluorobenzene	10.639	95	85892	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.584	96	3437	3.25	ug/l	96
16) Acetone	2.639	43	8059	6.92	ug/l	96
21) trans-1,2-Dichloroethene	3.359	96	1780	1.44	ug/l #	80
27) cis-1,2-Dichloroethene	4.674	96	18147	12.92	ug/l	92
40) Benzene	5.784	78	2333	0.42	ug/l #	88
44) Trichloroethene	6.549	130	64714	45.47	ug/l	93
52) Toluene	7.976	92	85100	24.46	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041621\
Data File : VU043166.D
Acq On : 16 Apr 2021 14:19
Operator : SY/MD
Sample : M2061-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GWBR1-225-250-041521

Quant Time: Apr 17 02:31:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

