

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041621\
 Data File : VU043174.D
 Acq On : 16 Apr 2021 17:24
 Operator : SY/MD
 Sample : M2040-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB-041321

Quant Time: Apr 17 02:33:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	111287	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	185831	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	182264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	85470	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	96828	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
35) Dibromofluoromethane	5.298	113	67175	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
50) Toluene-d8	7.906	98	237326	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
62) 4-Bromofluorobenzene	10.639	95	83484	42.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.38%	
Target Compounds						
16) Acetone	2.642	43	2843	2.28	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041621\
Data File : VU043174.D
Acq On : 16 Apr 2021 17:24
Operator : SY/MD
Sample : M2040-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB-041321

Quant Time: Apr 17 02:33:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

