

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041719\  
 Data File : VU031213.D  
 Acq On : 16 Apr 2019 16:08  
 Operator : JC/SP  
 Sample : VSTD00547  
 Misc : 5.0mL/MSVOA U/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VSTD00547

Manual Integrations  
 APPROVED

MMDadoda  
 4/17/2019 10:39:39 AM

Quant Time: Apr 17 02:42:54 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMULM041719WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Wed Apr 17 02:35:17 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.88  | 114  | 515296   | 50.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.09  | 117  | 478262   | 50.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.48 | 152  | 209927   | 50.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |       |     |        |      |      |      |
|--------------------------------|-------|-----|--------|------|------|------|
| 4) Vinyl Chloride-d3           | 1.40  | 65  | 24622  | 5.38 | ug/L | 0.00 |
| 7) Chloroethane-d5             | 1.68  | 69  | 20968  | 5.79 | ug/L | 0.00 |
| 11) 1,1-Dichloroethene-d2      | 2.27  | 63  | 51312  | 6.09 | ug/L | 0.00 |
| 21) 2-Butanone-d5              | 4.21  | 46  | 32233m | 8.91 | ug/L | 0.02 |
| 24) Chloroform-d               | 4.64  | 84  | 40532  | 5.61 | ug/L | 0.00 |
| 26) 1,2-Dichloroethane-d4      | 5.31  | 65  | 25051  | 5.19 | ug/L | 0.00 |
| 32) Benzene-d6                 | 5.34  | 84  | 76689  | 5.40 | ug/L | 0.00 |
| 36) 1,2-Dichloropropane-d6     | 6.33  | 67  | 24520  | 5.12 | ug/L | 0.00 |
| 41) Toluene-d8                 | 7.57  | 98  | 60882  | 4.86 | ug/L | 0.00 |
| 43) trans-1,3-Dichloropropene- | 7.85  | 79  | 11140  | 5.06 | ug/L | 0.00 |
| 47) 2-Hexanone-d5              | 8.33  | 63  | 19886  | 9.08 | ug/L | 0.02 |
| 57) 1,1,2,2-Tetrachloroethane- | 10.43 | 84  | 36600  | 5.67 | ug/L | 0.00 |
| 64) 1,2-Dichlorobenzene-d4     | 11.86 | 152 | 25611  | 6.09 | ug/L | 0.00 |

## Target Compounds

|                                |      |     |        |       | Qvalue    |
|--------------------------------|------|-----|--------|-------|-----------|
| 2) Dichlorodifluoromethane     | 1.20 | 85  | 27023  | 5.935 | ug/L 99   |
| 3) Chloromethane               | 1.32 | 50  | 33602  | 5.821 | ug/L 99   |
| 5) Vinyl chloride              | 1.40 | 62  | 32249  | 5.788 | ug/L 99   |
| 6) Bromomethane                | 1.63 | 94  | 21276  | 6.962 | ug/L 97   |
| 8) Chloroethane                | 1.70 | 64  | 20303  | 6.165 | ug/L 100  |
| 9) Trichlorofluoromethane      | 1.87 | 101 | 46184  | 6.809 | ug/L 97   |
| 10) 1,1,2-Trichloro-1,2,2-trif | 2.28 | 101 | 23070  | 6.161 | ug/L 95   |
| 12) 1,1-Dichloroethene         | 2.27 | 96  | 22610  | 6.290 | ug/L 94   |
| 13) Acetone                    | 2.34 | 43  | 28774  | 8.256 | ug/L 99   |
| 14) Carbon disulfide           | 2.47 | 76  | 60765  | 5.567 | ug/L 99   |
| 15) Methyl Acetate             | 2.62 | 43  | 24578  | 4.473 | ug/L 97   |
| 16) Methylene chloride         | 2.69 | 84  | 22003  | 5.212 | ug/L 98   |
| 17) trans-1,2-Dichloroethene   | 2.98 | 96  | 21444  | 5.795 | ug/L 96   |
| 18) Methyl tert-butyl Ether    | 2.99 | 73  | 62334  | 5.058 | ug/L 97   |
| 19) 1,1-Dichloroethane         | 3.44 | 63  | 41856  | 5.501 | ug/L 94   |
| 20) cis-1,2-Dichloroethene     | 4.22 | 96  | 23174  | 5.502 | ug/L 85   |
| 22) 2-Butanone                 | 4.30 | 43  | 37851m | 8.988 | ug/L      |
| 23) Bromochloromethane         | 4.55 | 128 | 11746  | 5.869 | ug/L 94   |
| 25) Chloroform                 | 4.67 | 83  | 39876  | 5.092 | ug/L 97   |
| 27) 1,2-Dichloroethane         | 5.40 | 62  | 30241  | 5.088 | ug/L 96   |
| 29) Cyclohexane                | 4.99 | 56  | 35512  | 5.182 | ug/L 97   |
| 30) 1,1,1-Trichloroethane      | 4.91 | 97  | 36005  | 6.080 | ug/L # 71 |
| 31) Carbon tetrachloride       | 5.13 | 117 | 30255  | 5.924 | ug/L 98   |
| 33) Benzene                    | 5.39 | 78  | 86496  | 5.501 | ug/L 100  |
| 34) Trichloroethene            | 6.18 | 95  | 22921  | 5.942 | ug/L 91   |
| 35) Methylcyclohexane          | 6.41 | 83  | 30980  | 5.257 | ug/L 97   |
| 37) 1,2-Dichloropropane        | 6.43 | 63  | 23015  | 5.259 | ug/L # 97 |
| 38) Bromodichloromethane       | 6.76 | 83  | 30032  | 5.617 | ug/L 95   |
| 39) cis-1,3-Dichloropropene    | 7.27 | 75  | 32243  | 5.162 | ug/L 98   |
| 40) 4-Methyl-2-pentanone       | 7.47 | 43  | 59043  | 8.599 | ug/L # 95 |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041719\  
 Data File : VU031213.D  
 Acq On : 16 Apr 2019 16:08  
 Operator : JC/SP  
 Sample : VSTD00547  
 Misc : 5.0mL/MSVOA U/WATER  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VSTD00547

Manual Integrations  
 APPROVED

MMDadoda  
 4/17/2019 10:39:39 AM

Quant Time: Apr 17 02:42:54 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMULM041719WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Wed Apr 17 02:35:17 2019  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|--------------------------------|-------|------|----------|-------|-------|----------|
| 42) Toluene                    | 7.64  | 91   | 80954    | 5.097 | ug/L  | 99       |
| 44) trans-1,3-Dichloropropene  | 7.88  | 75   | 27655    | 5.007 | ug/L  | 95       |
| 45) 1,1,2-Trichloroethane      | 8.07  | 97   | 20507    | 5.433 | ug/L  | 97       |
| 46) Tetrachloroethene          | 8.22  | 164  | 16092    | 5.674 | ug/L  | 97       |
| 48) 2-Hexanone                 | 8.38  | 43   | 47509    | 8.904 | ug/L  | 95       |
| 49) Dibromochloromethane       | 8.48  | 129  | 21620    | 5.389 | ug/L  | 99       |
| 50) 1,2-Dibromoethane          | 8.59  | 107  | 20562    | 5.072 | ug/L  | 95       |
| 51) Chlorobenzene              | 9.12  | 112  | 54607    | 5.560 | ug/L  | 95       |
| 52) Ethylbenzene               | 9.25  | 91   | 87611    | 5.055 | ug/L  | 99       |
| 53) m,p-Xylene                 | 9.37  | 106  | 26825    | 4.411 | ug/L  | 92       |
| 54) o-xylene                   | 9.78  | 106  | 29703    | 4.905 | ug/L  | 94       |
| 55) Styrene                    | 9.79  | 104  | 49827    | 4.820 | ug/L  | 96       |
| 56) Isopropylbenzene           | 10.16 | 105  | 78287    | 4.936 | ug/L  | 99       |
| 58) 1,1,2,2-Tetrachloroethane  | 10.46 | 83   | 36356    | 5.403 | ug/L  | 99       |
| 59) 1,2,3-Trichloropropane     | 10.49 | 75   | 28102    | 5.379 | ug/L  | 93       |
| 61) Bromoform                  | 9.96  | 173  | 15750    | 5.644 | ug/L  | 98       |
| 62) 1,3-Dichlorobenzene        | 11.42 | 146  | 39642    | 5.985 | ug/L  | 96       |
| 63) 1,4-Dichlorobenzene        | 11.51 | 146  | 39633    | 5.856 | ug/L  | 99       |
| 65) 1,2-Dichlorobenzene        | 11.88 | 146  | 41719    | 6.237 | ug/L  | 98       |
| 66) 1,2-Dibromo-3-chloropropan | 12.66 | 75   | 8017     | 5.787 | ug/L  | 86       |
| 67) 1,3,5-Trichlorobenzene     | 12.89 | 180  | 27416    | 5.623 | ug/L  | 97       |
| 68) 1,2,4-trichlorobenzene     | 13.50 | 180  | 19394    | 4.715 | ug/L  | 95       |
| 69) Naphthalene                | 13.75 | 128  | 59516    | 4.483 | ug/L  | 96       |
| 70) 1,2,3-Trichlorobenzene     | 13.99 | 180  | 23250    | 5.369 | ug/L  | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041719\  
Data File : VU031213.D  
Acq On : 16 Apr 2019 16:08  
Operator : JC/SP  
Sample : VSTD00547  
Misc : 5.0mL/MSVOA U/WATER  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VSTD00547

Manual Integrations  
APPROVED

MMDadoda  
4/17/2019 10:39:39 AM

Quant Time: Apr 17 02:42:54 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMULM041719WMA.M  
Quant Title : VOC Analysis  
QLast Update : Wed Apr 17 02:35:17 2019  
Response via : Initial Calibration

