

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041719\
 Data File : VU031235.D
 Acq On : 17 Apr 2019 01:49
 Operator : JC/SP
 Sample : K2339-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Apr 17 10:56:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM041719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 17 04:13:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	489503	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	478273	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	187834	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	169227	37.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.08%
7) Chloroethane-d5	1.68	69	163595	41.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.60%
11) 1,1-Dichloroethene-d2	2.27	63	280609	31.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.84%
21) 2-Butanone-d5	4.18	46	290477	88.47	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.47%
24) Chloroform-d	4.64	84	312113	41.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.98%
26) 1,2-Dichloroethane-d4	5.31	65	203421	43.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.84%
32) Benzene-d6	5.34	84	620146	41.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.04%
36) 1,2-Dichloropropane-d6	6.32	67	211037	43.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.02%
41) Toluene-d8	7.56	98	581884	43.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.48%
43) trans-1,3-Dichloropropene-	7.85	79	94142	40.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.28%
47) 2-Hexanone-d5	8.31	63	200037	86.35	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.35%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	280568	41.18	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.36%
64) 1,2-Dichlorobenzene-d4	11.86	152	188907	43.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.70%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041719\
Data File : VU031235.D
Acq On : 17 Apr 2019 01:49
Operator : JC/SP
Sample : K2339-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Apr 17 10:56:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM041719WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 17 04:13:40 2019
Response via : Initial Calibration

