

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041719\
 Data File : VU031246.D
 Acq On : 17 Apr 2019 09:49
 Operator : JC/SP
 Sample : VU0417MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK33

Quant Time: Apr 18 05:03:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM041719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 18 04:50:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	464490	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	455568	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	195348	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	246184	57.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	115.10%
7) Chloroethane-d5	1.68	69	201722	54.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.62%
11) 1,1-Dichloroethene-d2	2.27	63	337731	39.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.70%
21) 2-Butanone-d5	4.18	46	385425	123.71	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	123.71%
24) Chloroform-d	4.64	84	365473	51.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.42%
26) 1,2-Dichloroethane-d4	5.30	65	220941	50.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.54%
32) Benzene-d6	5.34	84	680529	47.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.50%
36) 1,2-Dichloropropane-d6	6.32	67	222869	47.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.36%
41) Toluene-d8	7.56	98	684271	53.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.98%
43) trans-1,3-Dichloropropene-	7.85	79	101366	45.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.74%
47) 2-Hexanone-d5	8.31	63	225569	102.23	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.23%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	328521	50.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.26%
64) 1,2-Dichlorobenzene-d4	11.85	152	223918	49.98	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.96%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041719\
Data File : VU031246.D
Acq On : 17 Apr 2019 09:49
Operator : JC/SP
Sample : VU0417MBL01
Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK33

Quant Time: Apr 18 05:03:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM041719WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 18 04:50:11 2019
Response via : Initial Calibration

