

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041723\
 Data File : VU053740.D
 Acq On : 17 Apr 2023 20:35
 Operator : JC/MD
 Sample : 02340-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHE16

Quant Time: Apr 18 00:23:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 18 00:19:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	139662	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	132020	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	64568	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	47226	4.915	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.400%
7) Chloroethane-d5	1.909	69	42584	4.884	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.600%
11) 1,1-Dichloroethene-d2	2.565	65	17675	4.319	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.400%
20) 2-Butanone-d5	4.623	46	97448	44.274	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.540%
24) Chloroform-d	5.057	84	84923	4.768	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
26) 1,2-Dichloroethane-d4	5.697	65	43717	4.789	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.800%
32) Benzene-d6	5.723	84	172250	4.604	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
36) 1,2-Dichloropropane-d6	6.684	67	55426	4.862	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.200%
41) Toluene-d8	7.893	98	155836	4.415	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
43) trans-1,3-Dichloroprop...	8.176	79	20344	4.650	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.000%
46) 2-Hexanone-d5	8.626	63	83766	47.138	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.280%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	42948	4.526	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	12.189	152	60914	5.244	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.800%
Target Compounds						
42) Toluene	7.967	91	17032	0.441	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041723\
Data File : VU053740.D
Acq On : 17 Apr 2023 20:35
Operator : JC/MD
Sample : 02340-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHE16

Quant Time: Apr 18 00:23:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 18 00:19:36 2023
Response via : Initial Calibration





#42
Toluene
Concen: 0.441 ug/L
RT: 7.967 min Scan# 2061
Delta R.T. 0.000 min
Lab File: VU053740.D
Acq: 17 Apr 2023 20:35

Instrument :
MSVOA_U
ClientSampleId :
BHE16

Tgt Ion: 91 Resp: 17032
Ion Ratio Lower Upper
91 100
92 55.1 40.4 75.0

