

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041819\
 Data File : VU031273.D
 Acq On : 17 Apr 2019 22:17
 Operator : JC/SP
 Sample : K2410-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0X36

Quant Time: Apr 18 09:04:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 18 08:52:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	253231	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	236677	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	94454	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	117916	5.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.80%
7) Chloroethane-d5	1.68	69	97735	5.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.80%
11) 1,1-Dichloroethene-d2	2.27	63	162424	3.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	4.21	46	274413	56.75	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.50%
24) Chloroform-d	4.64	84	154946	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
26) 1,2-Dichloroethane-d4	5.31	65	84085	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.40%
32) Benzene-d6	5.33	84	321279	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
36) 1,2-Dichloropropane-d6	6.33	67	106828	5.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.60%
41) Toluene-d8	7.56	98	273035	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
43) trans-1,3-Dichloropropene-	7.85	79	33352	4.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.80%
46) 2-Hexanone-d5	8.31	63	211746	52.91	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.82%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	86872	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	97057	5.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.60%

Target Compounds				Ovalue
13) Acetone	2.35	43	9747	2.470 ug/L 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041819\
Data File : VU031273.D
Acq On : 17 Apr 2019 22:17
Operator : JC/SP
Sample : K2410-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0X36

Quant Time: Apr 18 09:04:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 18 08:52:58 2019
Response via : Initial Calibration

