

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041819\
 Data File : VU031277.D
 Acq On : 17 Apr 2019 23:55
 Operator : JC/SP
 Sample : K2410-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0X40

Quant Time: Apr 18 09:29:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 18 08:39:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	240433	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	218485	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	90378	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	125550	6.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	126.60%
7) Chloroethane-d5	1.68	69	100524	6.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	122.20%
11) 1,1-Dichloroethene-d2	2.27	63	173095	4.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.60%
20) 2-Butanone-d5	4.22	46	280190	61.03	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.06%
24) Chloroform-d	4.64	84	161965	5.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.40%
26) 1,2-Dichloroethane-d4	5.31	65	88922	6.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	120.80%
32) Benzene-d6	5.33	84	334289	5.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.60%
36) 1,2-Dichloropropane-d6	6.33	67	111671	5.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	119.60%
41) Toluene-d8	7.56	98	279129	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
43) trans-1,3-Dichloropropene-	7.85	79	33327	4.43	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.60%
46) 2-Hexanone-d5	8.31	63	214572	58.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.16%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	88611	6.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	121.20%#
64) 1,2-Dichlorobenzene-d4	11.86	152	104866	6.19	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	123.80%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041819\
Data File : VU031277.D
Acq On : 17 Apr 2019 23:55
Operator : JC/SP
Sample : K2410-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0X40

Quant Time: Apr 18 09:29:36 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 18 08:39:25 2019
Response via : Initial Calibration

