

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041823\
 Data File : VU053762.D
 Acq On : 18 Apr 2023 16:14
 Operator : JC/MD
 Sample : VIBLK153
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK153

Quant Time: Apr 19 05:02:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 19 04:58:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	146421	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	136567	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	66153	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	41477	4.118	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.400%
7) Chloroethane-d5	1.912	69	39146	4.282	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.600%
11) 1,1-Dichloroethene-d2	2.565	65	15813	3.685	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.800%
20) 2-Butanone-d5	4.623	46	95730	41.485	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.980%
24) Chloroform-d	5.057	84	81425	4.360	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.200%
26) 1,2-Dichloroethane-d4	5.697	65	42891	4.482	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.600%
32) Benzene-d6	5.723	84	163045	4.213	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	6.687	67	52634	4.463	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.200%
41) Toluene-d8	7.896	98	143557	3.932	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.600%
43) trans-1,3-Dichloroprop...	8.176	79	18484	4.084	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%
46) 2-Hexanone-d5	8.629	63	78976	42.963	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.920%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	42212	4.300	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.000%
66) 1,2-Dichlorobenzene-d4	12.189	152	57071	4.795	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%
Target Compounds						Qvalue
16) Methylene chloride	3.044	84	3870	0.355	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041823\
Data File : VU053762.D
Acq On : 18 Apr 2023 16:14
Operator : JC/MD
Sample : VIBLK153
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK153

Quant Time: Apr 19 05:02:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 19 04:58:28 2023
Response via : Initial Calibration

