

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041919\
 Data File : VU031333.D
 Acq On : 19 Apr 2019 03:36
 Operator : JC/SP
 Sample : K2478-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0XD3

Quant Time: Apr 19 08:12:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 19 07:47:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	238628	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	244477	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	90347	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	90988	3.38	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.60%
7) Chloroethane-d5	1.68	69	74317	3.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.40%
11) 1,1-Dichloroethene-d2	2.27	63	134556	2.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.00%#
20) 2-Butanone-d5	4.20	46	294061	57.18	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.36%
24) Chloroform-d	4.64	84	173806	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.60%
26) 1,2-Dichloroethane-d4	5.31	65	98907	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.00%
32) Benzene-d6	5.34	84	327570	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	6.33	67	117747	5.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.00%
41) Toluene-d8	7.56	98	256582	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	7.85	79	37190	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
46) 2-Hexanone-d5	8.31	63	196382	49.53	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.06%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	95849	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	96486	5.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.60%


Target Compounds					Ovalue
13) Acetone	2.33	43	52723	9.214 ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041919\
Data File : VU031333.D
Acq On : 19 Apr 2019 03:36
Operator : JC/SP
Sample : K2478-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0XD3

Quant Time: Apr 19 08:12:13 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR041919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 19 07:47:43 2019
Response via : Initial Calibration





#13

Acetone

Concen: 9.214 ug/L

RT: 2.33 min Scan# 387

Delta R.T. -0.02 min

Lab File: VU031333.D

Acq: 19 Apr 2019 03:36

Instrument :
MSVOA_U
ClientSampleId :
C0XD3

Tot Ion: 43 Resp: 52723

Ion Ratio Lower Upper

43 100

58 32.5 0.0 64.4

