

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
 Data File : VU043186.D
 Acq On : 19 Apr 2021 12:05
 Operator : SY/MD
 Sample : VU0419WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0419WBL01

Quant Time: Apr 20 02:52:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	115805	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	193360	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.427	117	186356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.822	152	88177	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	98059	50.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.74%	
35) Dibromofluoromethane	5.298	113	68417	47.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.18%	
50) Toluene-d8	7.906	98	245929	47.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.66%	
62) 4-Bromofluorobenzene	10.642	95	86169	41.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.70%	
Target Compounds						
						Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
Data File : VU043186.D
Acq On : 19 Apr 2021 12:05
Operator : SY/MD
Sample : VU0419WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0419WBL01

Quant Time: Apr 20 02:52:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

