

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
 Data File : VU043191.D
 Acq On : 19 Apr 2021 14:08
 Operator : SY/MD
 Sample : M2077-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK-3

Quant Time: Apr 20 02:55:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	127558	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	205496	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	198051	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	94533	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	106113	49.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.96%	
35) Dibromofluoromethane	5.298	113	72725	47.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.20%	
50) Toluene-d8	7.906	98	262196	47.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.98%	
62) 4-Bromofluorobenzene	10.639	95	94236	43.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.14%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	11.475	105	2371	0.41	ug/l	98
95) Naphthalene	14.098	128	4785	3.40	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
Data File : VU043191.D
Acq On : 19 Apr 2021 14:08
Operator : SY/MD
Sample : M2077-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK-3

Quant Time: Apr 20 02:55:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

