

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
 Data File : VU043192.D
 Acq On : 19 Apr 2021 14:31
 Operator : SY/MD
 Sample : M2069-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-4-16-2021

Quant Time: Apr 20 02:55:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	122113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	199784	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	194396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	92683	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	101667	49.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.06%	
35) Dibromofluoromethane	5.298	113	70751	47.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.26%	
50) Toluene-d8	7.906	98	251528	47.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.70%	
62) 4-Bromofluorobenzene	10.639	95	91061	42.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.62%	
Target Compounds						
16) Acetone	2.642	43	1975	1.33	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
Data File : VU043192.D
Acq On : 19 Apr 2021 14:31
Operator : SY/MD
Sample : M2069-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-4-16-2021

Quant Time: Apr 20 02:55:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

