

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
 Data File : VU043193.D
 Acq On : 19 Apr 2021 14:54
 Operator : SY/MD
 Sample : M2078-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK-2

Quant Time: Apr 20 02:55:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 12 13:17:29 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	118573	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	196759	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	188569	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	90228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	101932	51.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.28%	
35) Dibromofluoromethane	5.298	113	70681	47.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.62%	
50) Toluene-d8	7.906	98	248890	47.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.14%	
62) 4-Bromofluorobenzene	10.639	95	90421	43.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041921\
Data File : VU043193.D
Acq On : 19 Apr 2021 14:54
Operator : SY/MD
Sample : M2078-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK-2

Quant Time: Apr 20 02:55:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U041221W.M
Quant Title : SW846 8260
QLast Update : Mon Apr 12 13:17:29 2021
Response via : Initial Calibration

